



File View Outputs System Help

Ramsey Error	-145	mV
Rabi Error	-2	mV
OCXO CV	-163	mV
Deviation	288	

Ramsey Error	415 mV
Rabi Error	-44 mV
C Field CV	0 mV

DC Level	1416	mV
DAC Gain	64.00	
Numeric Gain	4.00	
Clock Signal	2495	mV
Zeeman Signal	1895	mV

Attenuator CV -919 mV

Chassis Temp	39.30°C
Fault List	02,16,00,00,00
Power On Hours	0
Manufacture Date	00/00/00
Servo Firmware	

Delta-F E-15

OCXO Oven	Ready	
Cesium Oven	4.40	V
EM Voltage	9.20	V
Ion Pump	0	uA
C-Field	18.70	mA
Ionizer	0.76	V
Mass Spec	12.00	V

+24 VDC	26.48
+15 VDC	15.90
-15 VDC	-16.00
+5 VDC	5.10

Major Fault