

VERIFICATION OF COMPLIANCE

No.: LVD SZES1610004146MS

Applicant: Sanwa Electric Instrument Co., Ltd.
Dempa Building, 4-4 Sotokanda 2-Chome, Chiyoda-ku, Tokyo,
101-0021, Japan

Manufacturer: Sanwa Electric Instrument Co., Ltd.
4-7-15 Shinmeidai, Hamurashi, Tokyo 205-0023, Japan

Product Name: Digital Multimeter

Model No.: PM300

Rating: Max. measuring voltage: 600 V
Measurement category: CAT III 600 V, CAT IV 300 V

Protection against Electric Shock: Class II

Additional Information: --

Sufficient samples of the product have been tested and found to be in conformity with

Test Standard: EN 61010-1: 2010
EN 61010-2-030: 2010
EN 61010-2-033: 2012
EN 61010-031: 2015

as shown in the

Test Report Number(s): SZES161000414601

This Verification of Compliance has been granted to the applicant based on the results of tests, performed by Laboratory of SGS-CSTC Standards Technical Services Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant harmonized standards under the Low Voltage Directive 2014/35/EU. The CE marking as shown below can be affixed, under the responsibility of the manufacturer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives. The affixing of the CE marking presumes in addition that the conditions in annexes III and IV of the Directive are fulfilled.


Rocky Wang
Laboratory Manager
SGS-CSTC

 

2017-04-10

Copyright of this verification is owned by SGS-CSTC Standards Technical Services Co., Ltd. and may not be reproduced other than in full and with the prior approval of the General Manager. This verification is subjected to the governance of the General Conditions of Services, printed overleaf.

Member of SGS Group (Société Générale de Surveillance)